

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061618\
 Data File : VX002661.D
 Acq On : 16 Jun 2018 21:38
 Operator : JC/MD
 Sample : J3485-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-061318-03

Quant Time: Jun 18 11:16:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	207913	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	297481	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	279112	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	161000	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	141218	44.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.40%	
35) Dibromofluoromethane	5.51	113	108880	45.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.72%	
50) Toluene-d8	8.72	98	425578	47.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.42%	
62) 4-Bromofluorobenzene	11.14	95	150511	44.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.12%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061618\
Data File : VX002661.D
Acq On : 16 Jun 2018 21:38
Operator : JC/MD
Sample : J3485-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-061318-03

Quant Time: Jun 18 11:16:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

