

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061618\
 Data File : VX002660.D
 Acq On : 16 Jun 2018 21:11
 Operator : JC/MD
 Sample : J3462-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01_061218

Quant Time: Jun 18 11:16:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	213922	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	306964	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	286839	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	154887	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	146022	44.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.84%	
35) Dibromofluoromethane	5.51	113	111500	45.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.02%	
50) Toluene-d8	8.72	98	435055	46.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.54%	
62) 4-Bromofluorobenzene	11.14	95	148720	42.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.38%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061618\
Data File : VX002660.D
Acq On : 16 Jun 2018 21:11
Operator : JC/MD
Sample : J3462-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01_061218

Quant Time: Jun 18 11:16:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 833

Delta R.T. -0.01 min

Lab File: VX002660.D

Acq: 16 Jun 2018 21:11

Instrument :

MSVOA_X

ClientSampleId :


TB-01_061218

Tot Ion:168 Resp: 213922

Ion Ratio Lower Upper

168 100

99 50.8 39.3 58.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.67

80000



60000

40000

20000

0

Time--> 5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 44.919 ug/l

RT: 6.07 min Scan# 900

Delta R.T. 0.00 min

Lab File: VX002660.D


Acq: 16 Jun 2018 21:11

Tgt Ion: 65 Resp: 146022

Ion Ratio Lower Upper

65 100

67 52.0 0.0 102.6



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.07

60000

50000

40000


30000


20000

10000

0

Time--> 5.90 6.00 6.10 6.20





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.87 min Scan# 1030

Delta R.T. 0.00 min

Lab File: VX002660.D

Acq: 16 Jun 2018 21:11

Instrument :

MSVOA_X

ClientSampleId :

TB-01_061218


Tot Ion:114 Resp: 306964

Ion Ratio Lower Upper

114 100

63 20.0 0.0 41.4




88 14.7 0.0 30.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 45.513 ug/l

RT: 5.51 min Scan# 807

Delta R.T. 0.00 min

Lab File: VX002660.D

Acq: 16 Jun 2018 21:11


Tgt Ion:113 Resp: 111500

Ion Ratio Lower Upper

113 100

111 103.4 81.4 122.0



192 24.0 19.1 28.7




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 46.770 ug/l

RT: 8.72 min Scan# 1334

Delta R.T. 0.00 min

Lab File: VX002660.D

Acq: 16 Jun 2018 21:11

Instrument :

MSVOA_X

ClientSampleId :





TB-01_061218

Tot Ion: 98 Resp: 435055

Ion Ratio Lower Upper

98 100

100 63.7 50.9 76.3



#62

4-Bromofluorobenzene

Concen: 42.192 ug/l

RT: 11.14 min Scan# 1731

Delta R.T. 0.00 min

Lab File: VX002660.D

Acq: 16 Jun 2018 21:11




Tgt Ion: 95 Resp: 148720


Ion Ratio Lower Upper

95 100

174 97.3 0.0 187.4

176 91.9 0.0 181.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.12 min Scan# 1563
Delta R.T. 0.00 min
Lab File: VX002660.D
Acq: 16 Jun 2018 21:11

Instrument :
MSVOA_X
ClientSampleId :
TB-01_061218

Tot Ion:117 Resp: 286839
Ion Ratio Lower Upper
117 100
82 54.1 45.0 67.4
119 31.6 25.8 38.6



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.09 min Scan# 1886
Delta R.T. 0.00 min
Lab File: VX002660.D
Acq: 16 Jun 2018 21:11

Tgt Ion:152 Resp: 154887
Ion Ratio Lower Upper
152 100
115 54.7 40.1 120.2
150 157.1 0.0 347.2

