

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061818\
 Data File : VX002699.D
 Acq On : 19 Jun 2018 01:01
 Operator : JC/MD
 Sample : J3536-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z5TB08-180611

Quant Time: Jun 19 10:45:11 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	214582	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	300698	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	270513	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	141814	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	159464	48.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.80%	
35) Dibromofluoromethane	5.51	113	124210	51.76	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.52%	
50) Toluene-d8	8.72	98	473238	51.93	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.86%	
62) 4-Bromofluorobenzene	11.14	95	155748	45.11	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.22%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061818\
Data File : VX002699.D
Acq On : 19 Jun 2018 01:01
Operator : JC/MD
Sample : J3536-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z5TB08-180611

Quant Time: Jun 19 10:45:11 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

