

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061818\
 Data File : VX002701.D
 Acq On : 19 Jun 2018 01:54
 Operator : JC/MD
 Sample : J3536-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS050MW527-180611

Quant Time: Jun 19 10:46:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	218812	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	307958	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	287342	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	161290	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	163435	49.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.30%	
35) Dibromofluoromethane	5.51	113	127317	51.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.60%	
50) Toluene-d8	8.72	98	485523	52.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.06%	
62) 4-Bromofluorobenzene	11.14	95	168442	47.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.26%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	6241	2.313	ug/l	74
44) Trichloroethene	7.21	130	2775	0.980	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061818\
Data File : VX002701.D
Acq On : 19 Jun 2018 01:54
Operator : JC/MD
Sample : J3536-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS050MW527-180611

Quant Time: Jun 19 10:46:57 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

