

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061818\
 Data File : VX002705.D
 Acq On : 19 Jun 2018 03:40
 Operator : JC/MD
 Sample : J3536-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS050MW639-180612

Quant Time: Jun 19 07:06:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	206392	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	293599	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	273057	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	152271	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	153422	48.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.84%	
35) Dibromofluoromethane	5.51	113	120933	51.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.22%	
50) Toluene-d8	8.72	98	457266	51.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.80%	
62) 4-Bromofluorobenzene	11.14	95	156986	46.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.12%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	4165	1.636	ug/l	48
44) Trichloroethene	7.23	130	2241	0.830	ug/l	93
64) Tetrachloroethene	9.34	164	7144	2.357	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061818\
Data File : VX002705.D
Acq On : 19 Jun 2018 03:40
Operator : JC/MD
Sample : J3536-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS050MW639-180612

Quant Time: Jun 19 07:06:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

