

Data File : VX002711.D
Acq On : 19 Jun 2018 14:31
Operator : JC/MD
Sample : VX0619WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0619WBL01

Quant Time: Jun 19 16:23:58 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	202759	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	284390	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	257390	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	140120	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	147222	47.78	ug/l	0.00
Spiked Amount 50.000			Recovery =	95.56%		
35) Dibromofluoromethane	5.51	113	114967	50.65	ug/l	0.00
Spiked Amount 50.000			Recovery =	101.30%		
50) Toluene-d8	8.72	98	435495	50.53	ug/l	0.00
Spiked Amount 50.000			Recovery =	101.06%		
62) 4-Bromofluorobenzene	11.14	95	150785	46.17	ug/l	0.00
Spiked Amount 50.000			Recovery =	92.34%		

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX002711.D
Acq On : 19 Jun 2018 14:31
Operator : JC/MD
Sample : VX0619WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0619WBL01

Quant Time: Jun 19 16:23:58 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

