

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061918\
 Data File : VX002718.D
 Acq On : 19 Jun 2018 17:38
 Operator : JC/MD
 Sample : J3479-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z5TB07-180611

Quant Time: Jun 20 01:27:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	204170	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	287058	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	269808	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	152210	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	154279	49.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.46%	
35) Dibromofluoromethane	5.51	113	119482	52.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.30%	
50) Toluene-d8	8.72	98	457485	52.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.18%	
62) 4-Bromofluorobenzene	11.14	95	158703	48.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.30%	
Target Compounds						
16) Acetone	2.45	43	8773	6.41	ug/l	98
65) Chlorobenzene	10.14	112	1387	0.21	ug/l #	85
91) 1,2-Dichlorobenzene	12.40	146	3264	0.62	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061918\
Data File : VX002718.D
Acq On : 19 Jun 2018 17:38
Operator : JC/MD
Sample : J3479-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z5TB07-180611

Quant Time: Jun 20 01:27:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

