

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061918\
 Data File : VX002720.D
 Acq On : 19 Jun 2018 18:31
 Operator : JC/MD
 Sample : J3516-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IDW

Quant Time: Jun 20 17:02:37 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	213651	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	305333	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	282867	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	157668	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	162991	50.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.40%	
35) Dibromofluoromethane	5.51	113	127083	52.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.30%	
50) Toluene-d8	8.72	98	484531	52.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.74%	
62) 4-Bromofluorobenzene	11.14	95	167162	47.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.36%	
Target Compounds						
30) Chloroform	5.23	83	15637	3.48	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061918\
Data File : VX002720.D
Acq On : 19 Jun 2018 18:31
Operator : JC/MD
Sample : J3516-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IDW

Quant Time: Jun 20 17:02:37 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

