

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061918\
 Data File : VX002721.D
 Acq On : 19 Jun 2018 18:58
 Operator : JC/MD
 Sample : J3467-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 C-MW-10-061218

Quant Time: Jun 20 18:15:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	210691	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	301222	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	281601	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	162817	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	144114	45.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.02%	
35) Dibromofluoromethane	5.51	113	112471	46.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.56%	
50) Toluene-d8	8.72	98	429171	47.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.04%	
62) 4-Bromofluorobenzene	11.14	95	148964	43.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.14%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061918\
Data File : VX002721.D
Acq On : 19 Jun 2018 18:58
Operator : JC/MD
Sample : J3467-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C-MW-10-061218

Quant Time: Jun 20 18:15:05 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

