

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061918\
 Data File : VX002722.D
 Acq On : 19 Jun 2018 19:24
 Operator : JC/MD
 Sample : J3467-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 C-MW-06-061218

Quant Time: Jun 20 18:15:28 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	210892	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	299048	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	281296	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	157518	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	143242	44.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.40%	
35) Dibromofluoromethane	5.51	113	111714	46.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.62%	
50) Toluene-d8	8.72	98	427541	47.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.36%	
62) 4-Bromofluorobenzene	11.14	95	147196	42.86	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.72%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061918\
Data File : VX002722.D
Acq On : 19 Jun 2018 19:24
Operator : JC/MD
Sample : J3467-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C-MW-06-061218

Quant Time: Jun 20 18:15:28 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

