

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061918\
 Data File : VX002729.D
 Acq On : 19 Jun 2018 22:31
 Operator : JC/MD
 Sample : J3535-15
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS042MW024-180613

Quant Time: Jun 20 02:16:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	213093	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	300782	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	284659	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	162298	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	157432	48.62	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.24%	
35) Dibromofluoromethane	5.51	113	123984	51.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.30%	
50) Toluene-d8	8.72	98	464028	50.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.82%	
62) 4-Bromofluorobenzene	11.14	95	159826	46.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.54%	

Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	4.60	96	4016	1.53	ug/l	64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061918\
Data File : VX002729.D
Acq On : 19 Jun 2018 22:31
Operator : JC/MD
Sample : J3535-15
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS042MW024-180613

Quant Time: Jun 20 02:16:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

