

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061918\
 Data File : VX002738.D
 Acq On : 20 Jun 2018 03:22
 Operator : JC/MD
 Sample : J3517-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MANHATTANVILLE-FIELD-BLANK-1

Quant Time: Jun 20 07:27:52 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	225823	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	321262	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	301351	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	169511	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	152795	44.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.04%	
35) Dibromofluoromethane	5.51	113	119050	46.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.86%	
50) Toluene-d8	8.72	98	457490	46.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.98%	
62) 4-Bromofluorobenzene	11.14	95	156805	42.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.02%	
Target Compounds						
10) Methyl Iodide	2.53	142	2372	5.77	ug/l	93
16) Acetone	2.45	43	9257	6.11	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061918\
Data File : VX002738.D
Acq On : 20 Jun 2018 03:22
Operator : JC/MD
Sample : J3517-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MANHATTANVILLE-FIELD-BLANK-1

Quant Time: Jun 20 07:27:52 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

