

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061918\
 Data File : VX002741.D
 Acq On : 20 Jun 2018 04:54
 Operator : JC/MD
 Sample : J3510-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-22

Quant Time: Jun 20 07:39:46 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	221757	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	312361	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	295428	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	167029	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	150927	44.79	ug/l	0.00
Spiked Amount			Recovery	=	89.58%	
35) Dibromofluoromethane	5.51	113	117859	47.28	ug/l	0.00
Spiked Amount			Recovery	=	94.56%	
50) Toluene-d8	8.72	98	449734	47.51	ug/l	0.00
Spiked Amount			Recovery	=	95.02%	
62) 4-Bromofluorobenzene	11.14	95	156654	43.67	ug/l	0.00
Spiked Amount			Recovery	=	87.34%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.41	62	27239	9.37	ug/l #	81
16) Acetone	2.45	43	4630	3.11	ug/l	99
21) trans-1,2-Dichloroethene	3.17	96	29515	10.98	ug/l	98
27) cis-1,2-Dichloroethene	4.60	96	60821	22.24	ug/l	85
40) Benzene	6.15	78	3972	0.40	ug/l #	89
44) Trichloroethene	7.22	130	64128	22.34	ug/l	100
65) Chlorobenzene	10.14	112	28844	3.98	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061918\
Data File : VX002741.D
Acq On : 20 Jun 2018 04:54
Operator : JC/MD
Sample : J3510-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-22

Quant Time: Jun 20 07:39:46 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

