


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX061918\
Data File : VX002709.D
Acq On : 19 Jun 2018 13:07
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Title : SW846 8260
Last Update : Sat Jun 16 01:37:33 2018



AutoFind: Scans 1730, 1731, 1732; Background Corrected with Scan 1724

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.1	3631	PASS
75	95	30	60	54.1	9296	PASS
95	95	100	100	100.0	17190	PASS
96	95	5	9	7.3	1247	PASS
173	174	0.00	2	1.4	227	PASS
174	95	50	100	93.5	16078	PASS
175	174	5	9	8.3	1332	PASS
176	174	95	101	100.5	16154	PASS
177	176	5	9	6.9	1112	PASS