

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062018\
 Data File : VX002740.D
 Acq On : 20 Jun 2018 14:34
 Operator : JC/MD
 Sample : J3596-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-21

Quant Time: Jun 21 04:36:22 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	202709	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	284234	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	258729	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	142261	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	136404	44.28	ug/l	0.00
Spiked Amount			Recovery	=	88.56%	
35) Dibromofluoromethane	5.51	113	109352	48.21	ug/l	0.00
Spiked Amount			Recovery	=	96.42%	
50) Toluene-d8	8.72	98	399801	46.42	ug/l	0.00
Spiked Amount			Recovery	=	92.84%	
62) 4-Bromofluorobenzene	11.14	95	136136	41.71	ug/l	0.00
Spiked Amount			Recovery	=	83.42%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
10) Methyl Iodide	2.52	142	3008	6.01	ug/l	94
16) Acetone	2.45	43	6862	5.05	ug/l	96
30) Chloroform	5.22	83	339388	79.70	ug/l	99
44) Trichloroethene	7.21	130	1244	0.48	ug/l	77
47) Bromodichloromethane	7.90	83	17858	6.35	ug/l	92
60) Dibromochloromethane	9.59	129	9082	6.96	ug/l	96
71) Bromoform	10.86	173	11066	9.91	ug/l #	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062018\
Data File : VX002740.D
Acq On : 20 Jun 2018 14:34
Operator : JC/MD
Sample : J3596-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-21

Quant Time: Jun 21 04:36:22 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

