

Data File : VX002741.D
Acq On : 20 Jun 2018 15:01
Operator : JC/MD
Sample : J3596-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-22

Quant Time: Jul 02 05:17:04 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	216694	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	303768	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	286406	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	161017	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	148927	45.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.46%	
35) Dibromofluoromethane	5.51	113	118035	48.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.38%	
50) Toluene-d8	8.72	98	434199	47.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.34%	
62) 4-Bromofluorobenzene	11.14	95	150395	43.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.24%	
Target Compounds						
10) Methyl Iodide	2.52	142	1575	5.587	ug/l #	84
16) Acetone	2.45	43	3627	2.497	ug/l	92
30) Chloroform	5.21	83	1046	0.230	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX002741.D
Acq On : 20 Jun 2018 15:01
Operator : JC/MD
Sample : J3596-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-22

Quant Time: Jul 02 05:17:04 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

