

Data File : VX002743.D
Acq On : 20 Jun 2018 15:54
Operator : JC/MD
Sample : J3596-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-12

Quant Time: Jul 02 05:18:52 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	222239	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	310778	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	291187	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	167622	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	149569	44.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.58%	
35) Dibromofluoromethane	5.51	113	118599	47.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.64%	
50) Toluene-d8	8.72	98	444476	47.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.40%	
62) 4-Bromofluorobenzene	11.14	95	156010	43.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.44%	
Target Compounds						
					Qvalue	
10) Methyl Iodide	2.52	142	4058	6.200	ug/l #	96
16) Acetone	2.45	43	72148	48.422	ug/l	98
17) Carbon Disulfide	2.57	76	2960	0.503	ug/l	96
25) 2-Butanone	4.72	43	4328	2.212	ug/l #	78
30) Chloroform	5.22	83	1367	0.293	ug/l	91
51) 4-Methyl-2-Pentanone	8.66	43	1774	0.446	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX002743.D
Acq On : 20 Jun 2018 15:54
Operator : JC/MD
Sample : J3596-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-12

Quant Time: Jul 02 05:18:52 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

