

Data File : VX002744.D
Acq On : 20 Jun 2018 16:20
Operator : JC/MD
Sample : J3596-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-01

Quant Time: Jul 02 05:19:25 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	228574	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	323824	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	302111	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	178988	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	157112	45.23	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.46%	
35) Dibromofluoromethane	5.51	113	123026	47.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.20%	
50) Toluene-d8	8.72	98	462072	47.09	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.18%	
62) 4-Bromofluorobenzene	11.14	95	165206	44.43	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.86%	
Target Compounds						
10) Methyl Iodide	2.52	142	3146	5.949	ug/l	97
11) Tert butyl alcohol	3.08	59	2154	3.000	ug/l #	65
16) Acetone	2.45	43	36624	23.899	ug/l	97
25) 2-Butanone	4.71	43	5928	2.946	ug/l	99
51) 4-Methyl-2-Pentanone	8.65	43	9581	2.314	ug/l	96
84) 1,2,4-Trimethylbenzene	11.81	105	5251	0.509	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX002744.D
Acq On : 20 Jun 2018 16:20
Operator : JC/MD
Sample : J3596-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-01

Quant Time: Jul 02 05:19:25 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

