

Data File : VX002752.D
Acq On : 20 Jun 2018 19:53
Operator : JC/MD
Sample : J3597-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
W-23

Quant Time: Jul 02 06:33:47 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	220378	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	307930	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	284380	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	164881	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	146278	43.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.36%	
35) Dibromofluoromethane	5.51	113	120171	48.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.80%	
50) Toluene-d8	8.72	98	430017	46.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.16%	
62) 4-Bromofluorobenzene	11.14	95	153197	43.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.66%	
Target Compounds						
10) Methyl Iodide	2.51	142	1611	5.589	ug/l #	95
16) Acetone	2.45	43	1176	0.796	ug/l	91
19) Methyl tert-butyl Ether	3.20	73	5455	0.591	ug/l	98
65) Chlorobenzene	10.14	112	2958	0.424	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX002752.D
Acq On : 20 Jun 2018 19:53
Operator : JC/MD
Sample : J3597-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
W-23

Quant Time: Jul 02 06:33:47 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

