

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062018\
 Data File : VX002753.D
 Acq On : 20 Jun 2018 20:20
 Operator : JC/MD
 Sample : J3597-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 W-32

Quant Time: Jun 21 05:39:52 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	217110	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	300883	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	278356	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	159957	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	142195	43.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.20%	
35) Dibromofluoromethane	5.51	113	118293	49.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.52%	
50) Toluene-d8	8.72	98	424968	46.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.22%	
62) 4-Bromofluorobenzene	11.14	95	149269	43.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.40%	
Target Compounds						
10) Methyl Iodide	2.52	142	2204	5.75	ug/l	98
11) Tert butyl alcohol	3.08	59	1723	2.53	ug/l #	73
16) Acetone	2.45	43	1410	0.97	ug/l	91
19) Methyl tert-butyl Ether	3.21	73	18679	2.05	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062018\
Data File : VX002753.D
Acq On : 20 Jun 2018 20:20
Operator : JC/MD
Sample : J3597-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
W-32

Quant Time: Jun 21 05:39:52 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

