

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062018\
 Data File : VX002763.D
 Acq On : 21 Jun 2018 01:39
 Operator : JC/MD
 Sample : PB110408ZHE#02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB110408ZHE#02

Quant Time: Jun 21 06:59:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	219057	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	308774	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	290004	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	165891	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	149164	44.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.62%	
35) Dibromofluoromethane	5.51	113	119975	48.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.36%	
50) Toluene-d8	8.72	98	444088	47.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.92%	
62) 4-Bromofluorobenzene	11.14	95	153949	43.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.84%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
10) Methyl Iodide	2.53	142	2822	5.90	ug/l	96
16) Acetone	2.45	43	9632	6.56	ug/l	97
18) Methyl Acetate	2.78	43	22341	5.71	ug/l	96
20) Methylene Chloride	2.86	84	10754	3.57	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062018\
Data File : VX002763.D
Acq On : 21 Jun 2018 01:39
Operator : JC/MD
Sample : PB110408ZHE#02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB110408ZHE#02

Quant Time: Jun 21 06:59:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

