

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062018\
 Data File : VX002765.D
 Acq On : 21 Jun 2018 02:32
 Operator : JC/MD
 Sample : PB110408ZHE#04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB110408ZHE#04

Quant Time: Jun 21 07:03:37 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	226848	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	320121	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	298346	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	167272	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	154789	44.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.80%	
35) Dibromofluoromethane	5.51	113	123813	48.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.92%	
50) Toluene-d8	8.72	98	458395	47.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.50%	
62) 4-Bromofluorobenzene	11.14	95	158124	43.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.04%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
10) Methyl Iodide	2.52	142	1746	5.61	ug/l	89
16) Acetone	2.45	43	10444	6.87	ug/l	99
18) Methyl Acetate	2.78	43	23989	5.92	ug/l	98
20) Methylene Chloride	2.86	84	10548	3.38	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062018\
Data File : VX002765.D
Acq On : 21 Jun 2018 02:32
Operator : JC/MD
Sample : PB110408ZHE#04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB110408ZHE#04

Quant Time: Jun 21 07:03:37 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

