

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062018\
 Data File : VX002771.D
 Acq On : 21 Jun 2018 05:11
 Operator : JC/MD
 Sample : PB110408ZHE#10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB110408ZHE#10

Quant Time: Jun 21 07:32:56 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	218329	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	309435	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	287919	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	162063	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	150773	45.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.88%	
35) Dibromofluoromethane	5.51	113	120510	48.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.60%	
50) Toluene-d8	8.72	98	439416	46.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.72%	
62) 4-Bromofluorobenzene	11.14	95	154203	43.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.80%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	11172	7.63	ug/l	97
18) Methyl Acetate	2.78	43	23874	6.12	ug/l	97
20) Methylene Chloride	2.86	84	10823	3.61	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062018\
Data File : VX002771.D
Acq On : 21 Jun 2018 05:11
Operator : JC/MD
Sample : PB110408ZHE#10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB110408ZHE#10

Quant Time: Jun 21 07:32:56 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

