

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062018\
 Data File : VX002772.D
 Acq On : 21 Jun 2018 05:38
 Operator : JC/MD
 Sample : PB110408ZHE#11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB110408ZHE#11

Quant Time: Jun 21 07:45:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	217335	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	310051	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	290517	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	165926	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	150385	45.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.06%	
35) Dibromofluoromethane	5.51	113	120785	48.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.62%	
50) Toluene-d8	8.72	98	446589	47.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.06%	
62) 4-Bromofluorobenzene	11.14	95	156710	44.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.04%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	10726	7.36	ug/l	100
18) Methyl Acetate	2.78	43	23593	6.08	ug/l	94
20) Methylene Chloride	2.86	84	11596	3.88	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062018\
Data File : VX002772.D
Acq On : 21 Jun 2018 05:38
Operator : JC/MD
Sample : PB110408ZHE#11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB110408ZHE#11

Quant Time: Jun 21 07:45:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

