

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062018\
 Data File : VX002774.D
 Acq On : 21 Jun 2018 06:31
 Operator : JC/MD
 Sample : PB110408ZHE#13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB110408ZHE#13

Quant Time: Jun 21 07:53:50 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	232759	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	330681	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	311126	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	177627	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	161703	45.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.44%	
35) Dibromofluoromethane	5.51	113	129273	48.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.96%	
50) Toluene-d8	8.72	98	473448	47.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.50%	
62) 4-Bromofluorobenzene	11.14	95	165909	43.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.38%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	11092	7.11	ug/l	99
18) Methyl Acetate	2.78	43	23106	5.56	ug/l	98
20) Methylene Chloride	2.86	84	11584	3.62	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062018\
Data File : VX002774.D
Acq On : 21 Jun 2018 06:31
Operator : JC/MD
Sample : PB110408ZHE#13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB110408ZHE#13

Quant Time: Jun 21 07:53:50 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

