

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062018\
 Data File : VX002758.D
 Acq On : 20 Jun 2018 23:26
 Operator : JC/MD
 Sample : VX0620WBL02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0620WBL02

Quant Time: Jun 21 06:47:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	232409	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	325816	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	308261	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	173305	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	175485	49.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.38%	
35) Dibromofluoromethane	5.51	113	141210	54.30	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.60%	
50) Toluene-d8	8.72	98	513104	51.97	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.94%	
62) 4-Bromofluorobenzene	11.14	95	179376	47.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.88%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062018\
Data File : VX002758.D
Acq On : 20 Jun 2018 23:26
Operator : JC/MD
Sample : VX0620WBL02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0620WBL02

Quant Time: Jun 21 06:47:03 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 5.67 min Scan# 834

Delta R.T. 0.00 min

Lab File: VX002758.D

Acq: 20 Jun 2018 23:26

Instrument :

MSVOA_X

ClientSampleId :


VX0620WBL02

Tot Ion:168 Resp: 232409

Ion Ratio Lower Upper




168 100

99 47.7 39.3 58.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 49.69 ug/l

RT: 6.07 min Scan# 900

Delta R.T. 0.00 min

Lab File: VX002758.D


Acq: 20 Jun 2018 23:26

Tgt Ion: 65 Resp: 175485

Ion Ratio Lower Upper



65 100


67 51.0 0.0 102.6



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0





#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 6.87 min Scan# 1030

Delta R.T. 0.00 min

Lab File: VX002758.D

Acq: 20 Jun 2018 23:26

Instrument :

MSVOA_X

ClientSampleId :

VX0620WBL02


Tot Ion:114 Resp: 325816

Ion Ratio Lower Upper

114 100

63 20.4 0.0 41.4




88 15.3 0.0 30.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 54.30 ug/l

RT: 5.51 min Scan# 807

Delta R.T. 0.00 min

Lab File: VX002758.D

Acq: 20 Jun 2018 23:26


Tgt Ion:113 Resp: 141210

Ion Ratio Lower Upper

113 100

111 101.5 81.4 122.0



192 23.6 19.1 28.7




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 51.97 ug/l

RT: 8.72 min Scan# 1334

Delta R.T. 0.00 min

Lab File: VX002758.D

Acq: 20 Jun 2018 23:26

Instrument :

MSVOA_X

ClientSampleId :


VX0620WBL02

Tot Ion: 98 Resp: 513104

Ion Ratio Lower Upper

98 100

100 63.3 50.9 76.3



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX

8.72

300000



200000

100000

0

Time-->

8.60 8.70 8.80 8.90



#62

4-Bromofluorobenzene

Concen: 47.94 ug/l

RT: 11.14 min Scan# 1731

Delta R.T. 0.00 min

Lab File: VX002758.D

Acq: 20 Jun 2018 23:26


Tgt Ion: 95 Resp: 179376

Ion Ratio Lower Upper

95 100

174 97.2 0.0 187.4

176 94.3 0.0 181.0



Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V

200000

150000


100000


50000

0

Time-->

11.10 11.20









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.12 min Scan# 1563
Delta R.T. 0.00 min
Lab File: VX002758.D
Acq: 20 Jun 2018 23:26

Instrument :
MSVOA_X
ClientSampled :
VX0620WBL02

Tot Ion:117 Resp: 308261
Ion Ratio Lower Upper
117 100
82 54.1 45.0 67.4
119 32.5 25.8 38.6



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.09 min Scan# 1886
Delta R.T. 0.00 min
Lab File: VX002758.D
Acq: 20 Jun 2018 23:26

Tgt Ion:152 Resp: 173305
Ion Ratio Lower Upper
152 100
115 54.8 40.1 120.2
150 157.2 0.0 347.2

