

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062118\
 Data File : VX002787.D
 Acq On : 21 Jun 2018 14:43
 Operator : JC/MD
 Sample : J3617-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IDW-SO1-SRI3

Quant Time: Jun 22 01:33:32 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	237403	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	336624	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	318894	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	179898	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	163523	45.33	ug/l	0.00
Spiked Amount			Recovery	=	90.66%	
35) Dibromofluoromethane	5.51	113	129795	48.31	ug/l	0.00
Spiked Amount			Recovery	=	96.62%	
50) Toluene-d8	8.72	98	480351	47.09	ug/l	0.00
Spiked Amount			Recovery	=	94.18%	
62) 4-Bromofluorobenzene	11.14	95	169561	43.87	ug/l	0.00
Spiked Amount			Recovery	=	87.74%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
10) Methyl Iodide	2.51	142	1397	5.51	ug/l	# 86
16) Acetone	2.45	43	16199	10.18	ug/l	97
18) Methyl Acetate	2.78	43	18612	4.39	ug/l	100
20) Methylene Chloride	2.86	84	6166	1.89	ug/l	94
25) 2-Butanone	4.72	43	4245	2.03	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062118\
Data File : VX002787.D
Acq On : 21 Jun 2018 14:43
Operator : JC/MD
Sample : J3617-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IDW-SO1-SRI3

Quant Time: Jun 22 01:33:32 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

