

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX062118\
 Data File : VX002789.D
 Acq On : 21 Jun 2018 15:36
 Operator : JC/MD
 Sample : J3617-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SRI3-TB2

Quant Time: Jun 22 01:44:36 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	226555	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	320484	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	302769	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	174174	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	155676	45.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.44%	
35) Dibromofluoromethane	5.51	113	121198	47.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.76%	
50) Toluene-d8	8.72	98	455299	46.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.76%	
62) 4-Bromofluorobenzene	11.14	95	163257	44.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.72%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	144276	94.99	ug/l	100
18) Methyl Acetate	2.78	43	19421	4.80	ug/l	94
20) Methylene Chloride	2.86	84	7291	2.34	ug/l	97
25) 2-Butanone	4.71	43	97178	48.72	ug/l	99
51) 4-Methyl-2-Pentanone	8.65	43	2206	0.54	ug/l #	82
59) 2-Hexanone	9.50	43	34990	11.33	ug/l	100
65) Chlorobenzene	10.14	112	1546	0.21	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX062118\
Data File : VX002789.D
Acq On : 21 Jun 2018 15:36
Operator : JC/MD
Sample : J3617-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SRI3-TB2

Quant Time: Jun 22 01:44:36 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

