

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062118\
 Data File : VX002800.D
 Acq On : 21 Jun 2018 20:29
 Operator : JC/MD
 Sample : J3603-16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 26KV-TP-9

Quant Time: Jun 22 02:12:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	218448	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	310217	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	286945	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	163592	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	151845	45.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.48%	
35) Dibromofluoromethane	5.51	113	119350	48.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.42%	
50) Toluene-d8	8.72	98	440701	46.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.76%	
62) 4-Bromofluorobenzene	11.14	95	154584	43.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.80%	
Target Compounds						
16) Acetone	2.45	43	10368	7.08	ug/l	98
18) Methyl Acetate	2.78	43	19730	5.06	ug/l	97
20) Methylene Chloride	2.86	84	28040	9.34	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062118\
Data File : VX002800.D
Acq On : 21 Jun 2018 20:29
Operator : JC/MD
Sample : J3603-16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
26KV-TP-9

Quant Time: Jun 22 02:12:23 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

