

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062118\
 Data File : VX002805.D
 Acq On : 21 Jun 2018 23:35
 Operator : JC/MD
 Sample : VX0621WBL02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0621WBL02

Quant Time: Jun 22 02:33:27 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	233167	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	328486	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	309735	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	174250	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	160377	45.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.52%	
35) Dibromofluoromethane	5.51	113	126460	48.24	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.48%	
50) Toluene-d8	8.72	98	469634	47.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.36%	
62) 4-Bromofluorobenzene	11.14	95	164842	43.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.40%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062118\
Data File : VX002805.D
Acq On : 21 Jun 2018 23:35
Operator : JC/MD
Sample : VX0621WBL02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0621WBL02

Quant Time: Jun 22 02:33:27 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

