

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062118\
 Data File : VX002808.D
 Acq On : 22 Jun 2018 00:54
 Operator : JC/MD
 Sample : PB110441ZHE#11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB110441ZHE#11

Quant Time: Jun 22 03:09:54 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	243041	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	346074	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	322759	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	181106	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	166318	45.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.06%	
35) Dibromofluoromethane	5.51	113	130648	47.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.60%	
50) Toluene-d8	8.72	98	490263	46.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.50%	
62) 4-Bromofluorobenzene	11.14	95	171429	43.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.28%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	11263	6.91	ug/l	98
18) Methyl Acetate	2.78	43	21198	4.88	ug/l	99
20) Methylene Chloride	2.86	84	6346	1.90	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062118\
Data File : VX002808.D
Acq On : 22 Jun 2018 00:54
Operator : JC/MD
Sample : PB110441ZHE#11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB110441ZHE#11

Quant Time: Jun 22 03:09:54 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

