

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062118\
 Data File : VX002810.D
 Acq On : 22 Jun 2018 01:47
 Operator : JC/MD
 Sample : PB110441ZHE#13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB110441ZHE#13

Quant Time: Jun 22 03:16:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	233747	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	330561	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	310606	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	173964	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	163574	46.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.10%	
35) Dibromofluoromethane	5.51	113	126088	47.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.58%	
50) Toluene-d8	8.72	98	473238	47.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.48%	
62) 4-Bromofluorobenzene	11.14	95	162775	42.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.76%	
Target Compounds						
16) Acetone	2.45	43	13262	8.46	ug/l	99
18) Methyl Acetate	2.78	43	20651	4.95	ug/l	97
20) Methylene Chloride	2.86	84	6585	2.05	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062118\
Data File : VX002810.D
Acq On : 22 Jun 2018 01:47
Operator : JC/MD
Sample : PB110441ZHE#13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB110441ZHE#13

Quant Time: Jun 22 03:16:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

