

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062118\
 Data File : VX002813.D
 Acq On : 22 Jun 2018 03:07
 Operator : JC/MD
 Sample : PB110441ZHE#16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB110441ZHE#16

Quant Time: Jun 22 05:50:08 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	235490	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	334371	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	319914	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	182896	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	165558	46.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.52%	
35) Dibromofluoromethane	5.51	113	129044	48.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.72%	
50) Toluene-d8	8.72	98	482131	47.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.16%	
62) 4-Bromofluorobenzene	11.14	95	169603	44.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.34%	
Target Compounds						
16) Acetone	2.45	43	13099	8.30	ug/l	92
18) Methyl Acetate	2.78	43	22083	5.25	ug/l	97
20) Methylene Chloride	2.86	84	6095	1.88	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062118\
Data File : VX002813.D
Acq On : 22 Jun 2018 03:07
Operator : JC/MD
Sample : PB110441ZHE#16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB110441ZHE#16

Quant Time: Jun 22 05:50:08 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

