

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062118\
 Data File : VX002815.D
 Acq On : 22 Jun 2018 04:00
 Operator : JC/MD
 Sample : PB110441ZHE#18
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB110441ZHE#18

Quant Time: Jun 22 05:53:50 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	232566	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	329528	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	311864	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	175705	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	163711	46.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.64%	
35) Dibromofluoromethane	5.51	113	128169	48.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.46%	
50) Toluene-d8	8.72	98	474690	47.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.08%	
62) 4-Bromofluorobenzene	11.14	95	164487	43.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.94%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	13404	8.60	ug/l	93
18) Methyl Acetate	2.78	43	22775	5.48	ug/l	97
20) Methylene Chloride	2.85	84	6442	2.02	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062118\
Data File : VX002815.D
Acq On : 22 Jun 2018 04:00
Operator : JC/MD
Sample : PB110441ZHE#18
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB110441ZHE#18

Quant Time: Jun 22 05:53:50 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

