

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062118\
 Data File : VX002821.D
 Acq On : 22 Jun 2018 06:39
 Operator : JC/MD
 Sample : J3622-19
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SP-D

Quant Time: Jun 22 08:27:43 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	253634	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	360888	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	333426	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	193194	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	172585	44.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.56%	
35) Dibromofluoromethane	5.51	113	135576	47.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.14%	
50) Toluene-d8	8.72	98	508020	46.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.90%	
62) 4-Bromofluorobenzene	11.14	95	183932	44.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.76%	
Target Compounds						
16) Acetone	2.45	43	9614	5.65	ug/l	96
18) Methyl Acetate	2.78	43	21867	4.83	ug/l	96
20) Methylene Chloride	2.86	84	238171	68.31	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062118\
Data File : VX002821.D
Acq On : 22 Jun 2018 06:39
Operator : JC/MD
Sample : J3622-19
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SP-D

Quant Time: Jun 22 08:27:43 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

