

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062118\
 Data File : VX002822.D
 Acq On : 22 Jun 2018 07:05
 Operator : JC/MD
 Sample : J3622-15
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SP-C

Quant Time: Jun 22 08:29:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	254549	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	355977	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	338387	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	193652	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	174531	45.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.24%	
35) Dibromofluoromethane	5.51	113	136205	47.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.88%	
50) Toluene-d8	8.72	98	511149	47.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.76%	
62) 4-Bromofluorobenzene	11.14	95	182198	44.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.14%	
Target Compounds						
16) Acetone	2.45	43	10410	6.10	ug/l	99
18) Methyl Acetate	2.78	43	22464	4.94	ug/l	95
20) Methylene Chloride	2.86	84	477686	136.51	ug/l	97
84) 1,2,4-Trimethylbenzene	11.81	105	13259	1.19	ug/l	95
95) Naphthalene	13.84	128	68876	5.30	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062118\
Data File : VX002822.D
Acq On : 22 Jun 2018 07:05
Operator : JC/MD
Sample : J3622-15
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SP-C

Quant Time: Jun 22 08:29:26 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

