

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062118\
 Data File : VX002827.D
 Acq On : 22 Jun 2018 09:17
 Operator : JC/MD
 Sample : J3603-22
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 26KV-TP-6

Quant Time: Jun 22 09:45:21 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	242193	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	344789	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	325751	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	186780	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	167957	45.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.28%	
35) Dibromofluoromethane	5.51	113	132127	48.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.04%	
50) Toluene-d8	8.72	98	492814	47.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.34%	
62) 4-Bromofluorobenzene	11.14	95	174682	44.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.24%	
Target Compounds						
16) Acetone	2.45	43	13365	8.23	ug/l	96
18) Methyl Acetate	2.78	43	21095	4.88	ug/l	99
20) Methylene Chloride	2.86	84	66674	20.03	ug/l	99
95) Naphthalene	13.84	128	90424	7.21	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062118\
Data File : VX002827.D
Acq On : 22 Jun 2018 09:17
Operator : JC/MD
Sample : J3603-22
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
26KV-TP-6

Quant Time: Jun 22 09:45:21 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

