

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062118\
 Data File : VX002786.D
 Acq On : 21 Jun 2018 14:17
 Operator : JC/MD
 Sample : PB110456TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB110456TB

Quant Time: Jun 22 01:30:54 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	235429	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	329699	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	306543	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	166700	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	160647	44.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.80%	
35) Dibromofluoromethane	5.51	113	127584	48.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.98%	
50) Toluene-d8	8.72	98	468472	46.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.78%	
62) 4-Bromofluorobenzene	11.14	95	161873	42.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.52%	
Target Compounds						
16) Acetone	2.45	43	4824	3.06	ug/l	Qvalue 98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062118\
Data File : VX002786.D
Acq On : 21 Jun 2018 14:17
Operator : JC/MD
Sample : PB110456TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB110456TB

Quant Time: Jun 22 01:30:54 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 5.67 min Scan# 833

Delta R.T. -0.01 min

Lab File: VX002786.D

Acq: 21 Jun 2018 14:17

Instrument :

MSVOA_X

ClientSampleId :


PB110456TB

Tot Ion:168 Resp: 235429

Ion Ratio Lower Upper




168 100

99 50.7 39.3 58.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#16

Acetone

Concen: 3.06 ug/l

RT: 2.45 min Scan# 305

Delta R.T. -0.01 min

Lab File: VX002786.D


Acq: 21 Jun 2018 14:17

Tgt Ion: 43 Resp: 4824

Ion Ratio Lower Upper



43 100


58 28.8 22.1 33.1



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0









#33
 1,2-Dichloroethane-d4
 Concen: 44.90 ug/l
 RT: 6.07 min Scan# 900
 Delta R.T. 0.00 min
 Lab File: VX002786.D
 Acq: 21 Jun 2018 14:17




Instrument :
 MSVOA_X
 ClientSampleId :
 PB110456TB


Tot Ion: 65 Resp: 160647
 Ion Ratio Lower Upper
 65 100
 67 50.9 0.0 102.6



#34
 1,4-Difluorobenzene
 Concen: 50.00 ug/l
 RT: 6.87 min Scan# 1030
 Delta R.T. 0.00 min
 Lab File: VX002786.D
 Acq: 21 Jun 2018 14:17

Tgt Ion: 114 Resp: 329699
 Ion Ratio Lower Upper
 114 100
 63 20.0 0.0 41.4
 88 15.3 0.0 30.0





#35

Dibromofluoromethane

Concen: 48.49 ug/l

RT: 5.51 min Scan# 807

Delta R.T. 0.00 min

Lab File: VX002786.D

Acq: 21 Jun 2018 14:17

Instrument :

MSVOA_X

ClientSampleId :

PB110456TB


Tot Ion:113 Resp: 127584

Ion Ratio Lower Upper

113 100

111 100.3 81.4 122.0




192 24.1 19.1 28.7



Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V



#50

Toluene-d8

Concen: 46.89 ug/l

RT: 8.72 min Scan# 1334

Delta R.T. 0.00 min

Lab File: VX002786.D


Acq: 21 Jun 2018 14:17

Tgt Ion: 98 Resp: 468472

Ion Ratio Lower Upper



98 100


100 63.8 50.9 76.3



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): V





#62

4-Bromofluorobenzene

Concen: 42.76 ug/l

RT: 11.14 min Scan# 1731

Delta R.T. 0.00 min

Lab File: VX002786.D

Acq: 21 Jun 2018 14:17

Instrument :

MSVOA_X

ClientSampleId :

PB110456TB


Tot Ion: 95 Resp: 161873

Ion Ratio Lower Upper

95 100

174 96.7 0.0 187.4




176 94.6 0.0 181.0



Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V



#63

Chlorobenzene-d5

Concen: 50.00 ug/l

RT: 10.12 min Scan# 1563

Delta R.T. 0.00 min

Lab File: VX002786.D

Acq: 21 Jun 2018 14:17


Tgt Ion: 117 Resp: 306543

Ion Ratio Lower Upper

117 100

82 55.1 45.0 67.4



119 31.4 25.8 38.6






Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): VX0

Ion 118.90 (118.60 to 119.60): V





#72

1,4-Dichlorobenzene-d4

Concen: 50.00 ug/l

RT: 12.09 min Scan# 1886

Delta R.T. 0.00 min

Lab File: VX002786.D

Acq: 21 Jun 2018 14:17

Instrument :

MSVOA_X

ClientSampleId :

PB110456TB

Tot Ion:152 Resp: 166700

Ion Ratio Lower Upper

152 100

115 55.0 40.1 120.2

150 158.8 0.0 347.2

