


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX062118\
Data File : VX002781.D
Acq On : 21 Jun 2018 09:55
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Title : SW846 8260
Last Update : Sat Jun 16 01:37:33 2018



AutoFind: Scans 1730, 1731, 1732; Background Corrected with Scan 1723

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.7	4147	PASS
75	95	30	60	55.0	10515	PASS
95	95	100	100	100.0	19128	PASS
96	95	5	9	7.5	1441	PASS
173	174	0.00	2	1.5	278	PASS
174	95	50	100	99.6	19057	PASS
175	174	5	9	7.4	1414	PASS
176	174	95	101	95.4	18178	PASS
177	176	5	9	6.0	1090	PASS