

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062824\
 Data File : VX042094.D
 Acq On : 28 Jun 2024 18:35
 Operator : JC/MD
 Sample : P3128-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VPB189-HYD-20240626

Quant Time: Jun 29 04:23:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	146146	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	223373	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	222274	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	111472	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	76137	45.178	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.360%
35) Dibromofluoromethane	5.385	113	75315	47.519	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.040%
50) Toluene-d8	8.647	98	256999	46.156	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.320%
62) 4-Bromofluorobenzene	11.079	95	106748	52.009	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.020%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	1932	2.611	ug/l	92
44) Trichloroethene	7.135	130	729	0.479	ug/l	67
60) Dibromochloromethane	9.525	129	615	0.376	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062824\
Data File : VX042094.D
Acq On : 28 Jun 2024 18:35
Operator : JC/MD
Sample : P3128-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VPB189-HYD-20240626

Quant Time: Jun 29 04:23:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

