

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062824\
 Data File : VX042095.D
 Acq On : 28 Jun 2024 18:58
 Operator : JC/MD
 Sample : P3128-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-189-GW-625-627

Quant Time: Jun 29 04:24:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	138891	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	212578	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	209881	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	100783	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	73800	46.079	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.160%
35) Dibromofluoromethane	5.385	113	72237	47.891	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.780%
50) Toluene-d8	8.647	98	269021	50.769	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.540%
62) 4-Bromofluorobenzene	11.079	95	98197	50.272	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.540%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.319	101	24564	20.454	ug/l	# 85
12) 1,1-Dichloroethene	2.313	96	3454	2.883	ug/l	90
16) Acetone	2.392	43	2761	3.926	ug/l	93
27) cis-1,2-Dichloroethene	4.495	96	6482	4.257	ug/l	76
38) Carbon Tetrachloride	5.666	117	1270	0.718	ug/l	87
44) Trichloroethene	7.123	130	275041	189.943	ug/l	86
64) Tetrachloroethene	9.281	164	1151	0.753	ug/l	# 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062824\
Data File : VX042095.D
Acq On : 28 Jun 2024 18:58
Operator : JC/MD
Sample : P3128-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-189-GW-625-627

Quant Time: Jun 29 04:24:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

