

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX062824\  
 Data File : VX042096.D  
 Acq On : 28 Jun 2024 19:22  
 Operator : JC/MD  
 Sample : P3128-06  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 17 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 BP-VPB-189-GW-638-640

Quant Time: Jun 29 04:24:31 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X061824W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Jun 21 23:51:58 2024  
 Response via : Initial Calibration

| Compound                     | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|------------------------------|--------|-------|----------|----------|-------|----------|
| -----                        |        |       |          |          |       |          |
| Internal Standards           |        |       |          |          |       |          |
| 1) Pentafluorobenzene        | 5.556  | 168   | 138748   | 50.000   | ug/l  | # 0.00   |
| 34) 1,4-Difluorobenzene      | 6.763  | 114   | 212522   | 50.000   | ug/l  | 0.00     |
| 63) Chlorobenzene-d5         | 10.055 | 117   | 207380   | 50.000   | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 12.024 | 152   | 98882    | 50.000   | ug/l  | 0.00     |
| System Monitoring Compounds  |        |       |          |          |       |          |
| 33) 1,2-Dichloroethane-d4    | 5.958  | 65    | 72536    | 45.337   | ug/l  | 0.00     |
| Spiked Amount                | 50.000 | Range | 78 - 117 | Recovery | =     | 90.680%  |
| 35) Dibromofluoromethane     | 5.385  | 113   | 72228    | 47.898   | ug/l  | 0.00     |
| Spiked Amount                | 50.000 | Range | 75 - 124 | Recovery | =     | 95.800%  |
| 50) Toluene-d8               | 8.647  | 98    | 263285   | 49.700   | ug/l  | 0.00     |
| Spiked Amount                | 50.000 | Range | 92 - 112 | Recovery | =     | 99.400%  |
| 62) 4-Bromofluorobenzene     | 11.079 | 95    | 96595    | 49.465   | ug/l  | 0.00     |
| Spiked Amount                | 50.000 | Range | 83 - 123 | Recovery | =     | 98.920%  |
| Target Compounds             |        |       |          |          |       |          |
|                              |        |       |          |          |       | Qvalue   |
| 9) 1,1,2-Trichlorotrifluo... | 2.319  | 101   | 22551    | 18.797   | ug/l  | # 85     |
| 12) 1,1-Dichloroethene       | 2.313  | 96    | 2431     | 2.031    | ug/l  | # 76     |
| 16) Acetone                  | 2.392  | 43    | 2111     | 3.005    | ug/l  | 92       |
| 27) cis-1,2-Dichloroethene   | 4.489  | 96    | 5932     | 3.900    | ug/l  | 78       |
| 44) Trichloroethene          | 7.123  | 130   | 168796   | 116.601  | ug/l  | 86       |
| 64) Tetrachloroethene        | 9.275  | 164   | 1147     | 0.760    | ug/l  | 91       |
| -----                        |        |       |          |          |       |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX062824\  
Data File : VX042096.D  
Acq On : 28 Jun 2024 19:22  
Operator : JC/MD  
Sample : P3128-06  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 17 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
BP-VPB-189-GW-638-640

Quant Time: Jun 29 04:24:31 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X061824W.M  
Quant Title : SW846 8260  
QLast Update : Fri Jun 21 23:51:58 2024  
Response via : Initial Calibration

