

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062824\
 Data File : VX042097.D
 Acq On : 28 Jun 2024 19:45
 Operator : JC/MD
 Sample : P3128-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-189-DUP-20240627

Quant Time: Jun 29 04:24:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	136319	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	208201	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	206250	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95675	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	71373	45.405	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.800%
35) Dibromofluoromethane	5.385	113	70243	47.548	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.100%
50) Toluene-d8	8.647	98	257501	49.616	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.240%
62) 4-Bromofluorobenzene	11.079	95	94091	49.183	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.360%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.319	101	38442	32.614	ug/l	# 84
12) 1,1-Dichloroethene	2.306	96	1092	0.929	ug/l	81
16) Acetone	2.386	43	2452	3.552	ug/l	# 87
27) cis-1,2-Dichloroethene	4.495	96	4404	2.947	ug/l	73
44) Trichloroethene	7.123	130	192465	135.711	ug/l	87
64) Tetrachloroethene	9.275	164	9251	6.160	ug/l	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062824\
Data File : VX042097.D
Acq On : 28 Jun 2024 19:45
Operator : JC/MD
Sample : P3128-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-189-DUP-20240627

Quant Time: Jun 29 04:24:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

