

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070722\
 Data File : VX029955.D
 Acq On : 07 Jul 2022 15:17
 Operator : JC/MD
 Sample : N3617-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-2022-07-06

Quant Time: Jul 08 05:34:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	197115	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	340015	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	292550	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	123114	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	106642	42.740	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	85.480%
35) Dibromofluoromethane	5.391	113	98906	46.546	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.100%
50) Toluene-d8	8.653	98	385167	48.318	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.640%
62) 4-Bromofluorobenzene	11.085	95	127492	45.245	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.480%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	4718	3.865	ug/l	89
20) Methylene Chloride	2.788	84	1187	0.460	ug/l #	90
44) Trichloroethene	7.135	130	822	0.334	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070722\
Data File : VX029955.D
Acq On : 07 Jul 2022 15:17
Operator : JC/MD
Sample : N3617-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-2022-07-06

Quant Time: Jul 08 05:34:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

