

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070722\
 Data File : VX029956.D
 Acq On : 07 Jul 2022 15:40
 Operator : JC/MD
 Sample : N3617-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-01

Quant Time: Jul 08 05:34:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	197394	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	333278	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	289605	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	123405	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	106358	42.566	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	85.140%
35) Dibromofluoromethane	5.391	113	98087	47.094	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.180%
50) Toluene-d8	8.653	98	383445	49.074	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.140%
62) 4-Bromofluorobenzene	11.086	95	123844	44.839	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.680%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	2978	2.436	ug/l	93
20) Methylene Chloride	2.788	84	671	0.260	ug/l #	77
44) Trichloroethene	7.135	130	1195	0.495	ug/l	65
65) Chlorobenzene	10.080	112	5698	0.980	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070722\
Data File : VX029956.D
Acq On : 07 Jul 2022 15:40
Operator : JC/MD
Sample : N3617-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-01

Quant Time: Jul 08 05:34:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

