

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070722\
 Data File : VX029957.D
 Acq On : 07 Jul 2022 16:04
 Operator : JC/MD
 Sample : N3617-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2R

Quant Time: Jul 08 05:34:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	193381	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	334179	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	292907	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	129707	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	107093	43.749	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.500%
35) Dibromofluoromethane	5.397	113	95882	45.911	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.820%
50) Toluene-d8	8.653	98	380929	48.620	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.240%
62) 4-Bromofluorobenzene	11.086	95	128536	46.412	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.820%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	5308	4.432	ug/l	92
20) Methylene Chloride	2.788	84	876	0.346	ug/l	87
44) Trichloroethene	7.135	130	773	0.320	ug/l	93
67) Ethyl Benzene	10.195	91	4287	0.419	ug/l	92
69) o-Xylene	10.640	106	861	0.217	ug/l	82
73) Isopropylbenzene	10.964	105	10641	1.134	ug/l	96
78) n-propylbenzene	11.311	91	7781	0.703	ug/l	98
85) sec-Butylbenzene	11.890	105	7982	0.820	ug/l	100
95) Naphthalene	13.780	128	22939	2.190	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070722\
Data File : VX029957.D
Acq On : 07 Jul 2022 16:04
Operator : JC/MD
Sample : N3617-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2R

Quant Time: Jul 08 05:34:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

