

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070722\
 Data File : VX029958.D
 Acq On : 07 Jul 2022 16:27
 Operator : JC/MD
 Sample : N3617-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-03

Quant Time: Jul 08 05:34:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	194225	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	335660	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	295359	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	128368	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	106856	43.463	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.920%
35) Dibromofluoromethane	5.391	113	96890	46.189	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.380%
50) Toluene-d8	8.653	98	381516	48.481	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.960%
62) 4-Bromofluorobenzene	11.085	95	130601	46.949	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.900%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	3984	3.312	ug/l	95
19) Methyl tert-butyl Ether	3.123	73	2152	0.319	ug/l #	81
20) Methylene Chloride	2.788	84	988	0.389	ug/l #	82
44) Trichloroethene	7.135	130	934	0.384	ug/l	93
83) tert-Butylbenzene	11.713	119	2645	0.346	ug/l	74
85) sec-Butylbenzene	11.896	105	5366	0.557	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070722\
Data File : VX029958.D
Acq On : 07 Jul 2022 16:27
Operator : JC/MD
Sample : N3617-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-03

Quant Time: Jul 08 05:34:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

