

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070722\
 Data File : VX029960.D
 Acq On : 07 Jul 2022 17:14
 Operator : JC/MD
 Sample : N3617-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-05

Quant Time: Jul 08 05:35:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	192143	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	338381	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	290614	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	126849	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	107056	44.016	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.040%
35) Dibromofluoromethane	5.391	113	97883	46.287	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.580%
50) Toluene-d8	8.653	98	380633	47.979	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.960%
62) 4-Bromofluorobenzene	11.085	95	127454	45.450	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.900%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	3216	2.703	ug/l #	85
19) Methyl tert-butyl Ether	3.117	73	2191	0.328	ug/l	94
44) Trichloroethene	7.135	130	843	0.344	ug/l	86
73) Isopropylbenzene	10.970	105	3630	0.395	ug/l	88
83) tert-Butylbenzene	11.719	119	2281	0.302	ug/l	80
85) sec-Butylbenzene	11.890	105	8594	0.903	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070722\
Data File : VX029960.D
Acq On : 07 Jul 2022 17:14
Operator : JC/MD
Sample : N3617-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-05

Quant Time: Jul 08 05:35:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

