

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070722\
 Data File : VX029962.D
 Acq On : 07 Jul 2022 18:01
 Operator : JC/MD
 Sample : N3617-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-07

Quant Time: Jul 08 05:35:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	185463	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	316658	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	274525	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	117688	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	102379	43.609	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.220%
35) Dibromofluoromethane	5.391	113	91579	46.277	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.560%
50) Toluene-d8	8.653	98	359841	48.470	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.940%
62) 4-Bromofluorobenzene	11.085	95	120009	45.731	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.460%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	6802	5.922	ug/l	98
44) Trichloroethene	7.141	130	811	0.354	ug/l	69
67) Ethyl Benzene	10.201	91	4230	0.441	ug/l #	88
73) Isopropylbenzene	10.963	105	11068	1.300	ug/l	100
78) n-propylbenzene	11.305	91	7933	0.790	ug/l	98
85) sec-Butylbenzene	11.890	105	7873	0.892	ug/l #	77
95) Naphthalene	13.774	128	21225	2.233	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070722\
Data File : VX029962.D
Acq On : 07 Jul 2022 18:01
Operator : JC/MD
Sample : N3617-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-07

Quant Time: Jul 08 05:35:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

