

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071825\
 Data File : VX047049.D
 Acq On : 18 Jul 2025 15:00
 Operator : JC/MD
 Sample : Q2637-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Jul 19 04:25:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 03 05:51:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	448042	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	792920	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	740369	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	367208	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	318516	53.812	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	107.620%
35) Dibromofluoromethane	5.397	113	264566	49.154	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.300%
50) Toluene-d8	8.653	98	953764	50.224	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.440%
62) 4-Bromofluorobenzene	11.079	95	380575	52.731	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071825\
Data File : VX047049.D
Acq On : 18 Jul 2025 15:00
Operator : JC/MD
Sample : Q2637-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Jul 19 04:25:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 03 05:51:11 2025
Response via : Initial Calibration

